
Midmark M11 - v1.0.0
Total Cycles: 2114

Name: RITTER 1
Sterilizer ID: V2178198
07/11/2022 01:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.5 F 0.0
2:00 151.9 F 1.5
4:00 172.2 F 4.6
6:00 187.2 F 8.7
8:00 211.9 F 13.3
10:00 225.2 F 7.0
12:00 235.2 F 11.8
14:00 233.3 F 8.4
16:00 243.0 F 11.7
18:00 243.9 F 12.4
20:00 256.1 F 18.9
22:00 267.1 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.6 F 29.5
2:00 272.3 F 28.9
3:00 271.4 F 28.7
4:00 272.5 F 29.5

Min 271.4 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 219.0 F 0.0
10:00 227.1 F 0.0
15:00 221.8 F 0.0
20:00 223.2 F 0.0
25:00 218.3 F 0.0
30:00 218.7 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 23:34
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:41
CYCLE COMPLETE

07/11/2022 02:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____