

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 007532  
6:38 12/06/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:31  
PRESSURIZING 3:32  
STERILIZING 4:40  
136.4 C 222kPa 4:40  
135.8 C 220kPa 5:10  
136.4 C 221kPa 5:40  
136.2 C 219kPa 6:10  
136.4 C 220kPa 6:40  
136.3 C 221kPa 7:10  
136.1 C 221kPa 7:40  
136.3 C 222kPa 8:10  
136.3 C 222kPa 8:10  
VENTING 8:10  
AIR DRYING 9:08  
CYCLE COMPLETE 39:08

-----  
Digital Signature #  
EEE55E2CF1DF3F99  
-----