

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2738

Name: RITTER2  
Sterilizer ID: V2313228  
08/08/2023 10:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.4 F 0.0  
2:00 86.6 F 0.0  
4:00 101.1 F 0.0  
6:00 113.0 F 0.0  
8:00 152.6 F 0.0  
10:00 180.1 F 0.7  
12:00 195.1 F 4.4  
14:00 212.9 F 7.5  
16:00 219.5 F 11.2  
18:00 224.5 F 6.3  
20:00 236.5 F 10.4  
22:00 235.0 F 8.5  
24:00 245.0 F 13.2  
26:00 242.8 F 11.9  
28:00 255.1 F 18.3  
30:00 265.9 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 28.7  
1:00 272.7 F 29.3  
2:00 271.9 F 29.1  
3:00 272.0 F 28.7  
4:00 272.7 F 29.4

Min 271.5 F 28.5  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.7 F 0.0  
5:00 230.0 F 0.0  
10:00 204.7 F 0.0  
15:00 201.0 F 0.0  
20:00 204.9 F 0.0  
25:00 209.6 F 0.0  
30:00 211.6 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 31:34  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 30:00  
TOTAL CYCLE: 01:09:25  
CYCLE COMPLETE

08/08/2023 11:49 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_