
Midmark M11 - v1.0.5
Total Cycles: 3097

Name: RITTER2
Sterilizer ID: V2313228
11/08/2023 09:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.1 F 0.0
2:00 80.2 F 0.0
4:00 120.6 F 0.2
6:00 108.8 F 1.5
8:00 141.2 F 4.3
10:00 169.8 F 7.4
12:00 207.0 F 11.5
14:00 214.5 F 4.0
16:00 224.7 F 7.8
18:00 238.3 F 11.9
20:00 231.6 F 7.7
22:00 243.0 F 12.9
24:00 240.4 F 10.9
26:00 252.5 F 17.0
28:00 263.6 F 23.3
30:00 272.5 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.8
1:00 272.6 F 29.1
2:00 272.1 F 29.5
3:00 271.4 F 28.7
4:00 272.4 F 29.4

Min 271.0 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.9 F 0.0
5:00 221.5 F 0.0
10:00 197.2 F 0.0
15:00 189.5 F 0.0
20:00 183.9 F 0.0
25:00 180.9 F 0.0
30:00 179.0 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 30:14
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:08:12
CYCLE COMPLETE

11/08/2023 10:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____