

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1084

Name: RITTER2  
Sterilizer ID: V2313228  
05/09/2022 12:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 137.4 F 0.0  
2:00 138.0 F 0.1  
4:00 154.5 F 3.4  
6:00 181.9 F 7.6  
8:00 220.1 F 12.2  
10:00 228.5 F 7.2  
12:00 241.2 F 12.3  
14:00 236.3 F 9.2  
16:00 238.3 F 10.2  
18:00 246.9 F 13.8  
20:00 259.0 F 20.4  
22:00 269.5 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.1  
1:00 272.4 F 29.2  
2:00 272.4 F 29.5  
3:00 272.0 F 29.6  
4:00 271.8 F 29.1

Min 271.6 F 29.0  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.0 F 0.0  
5:00 185.3 F 0.0  
10:00 193.2 F 0.0  
15:00 190.4 F 0.0  
20:00 184.8 F 0.0  
25:00 182.6 F 0.0  
30:00 180.0 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 23:49  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:01:59  
CYCLE COMPLETE

05/09/2022 01:12 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02