
Midmark M11 - v1.0.0
Total Cycles: 2872

Name: RITTER 1
Sterilizer ID: V2178198
02/08/2023 08:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.7 F 0.0
2:00 82.0 F 0.1
4:00 88.1 F 1.2
6:00 107.9 F 3.5
8:00 142.0 F 6.3
10:00 174.1 F 9.9
12:00 195.0 F 9.3
14:00 208.4 F 4.8
16:00 221.1 F 8.5
18:00 231.3 F 12.6
20:00 233.2 F 8.5
22:00 243.2 F 12.3
24:00 243.0 F 11.8
26:00 254.8 F 18.0
28:00 265.8 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.5
1:00 272.4 F 29.3
2:00 272.0 F 28.7
3:00 271.3 F 28.7
4:00 272.4 F 29.2

Min 271.1 F 28.3
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.6 F 0.0
5:00 204.2 F 0.0
10:00 216.1 F 0.0
15:00 223.5 F 0.0
20:00 226.2 F 0.0
25:00 230.2 F 0.0
30:00 230.2 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 29:50
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:07:52
CYCLE COMPLETE

02/08/2023 09:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____