
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001274
10:07 27/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:39
STERILIZING 3:50
136.3 C 223kPa 3:50
136.4 C 225kPa 4:20
136.4 C 223kPa 4:50
136.0 C 222kPa 5:20
136.4 C 225kPa 5:50
136.4 C 225kPa 6:20
136.3 C 222kPa 6:50
136.1 C 222kPa 7:20
136.1 C 222kPa 7:20
VENTING 7:20
AIR DRYING 8:12
CYCLE COMPLETE 13:36

Digital Signature #
D12C4863AC09143C
