
Midmark M11 - v1.0.0
Total Cycles: 1900

Name: RITTER 1
Sterilizer ID: V2178198
05/03/2022 09:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.9 F 0.0
2:00 189.2 F 3.3
4:00 196.6 F 7.8
6:00 217.1 F 13.1
8:00 227.5 F 8.6
10:00 237.3 F 10.0
12:00 241.3 F 11.8
14:00 239.6 F 10.0
16:00 253.3 F 16.9
18:00 264.9 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.5 F 29.1
2:00 272.0 F 28.5
3:00 271.2 F 28.4
4:00 272.4 F 29.2

Min 271.1 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.1 F 0.0
5:00 220.4 F 0.0
10:00 238.2 F 0.0
15:00 233.6 F 0.0
20:00 228.7 F 0.0
25:00 229.3 F 0.0
30:00 230.2 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 19:53
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:25
CYCLE COMPLETE

05/03/2022 10:53 AM
ACCEPTED: X
REJECTED:
Operator: U02