
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002230
8:44 07/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:10
STERILIZING 4:13
136.4 C 223kPa 4:13
136.4 C 224kPa 4:43
136.1 C 221kPa 5:13
136.3 C 223kPa 5:43
136.4 C 225kPa 6:13
136.4 C 225kPa 6:43
136.4 C 224kPa 7:13
136.4 C 223kPa 7:43
136.4 C 223kPa 7:43
VENTING 7:43
AIR DRYING 8:32
CYCLE COMPLETE 38:32

Digital Signature #
EDC428D156F7811A
