
Midmark M11 - v1.0.5
Total Cycles: 1581

Name: RITTER2
Sterilizer ID: V2313228
10/12/2022 07:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 88.2 F 0.0
2:00 125.6 F 0.0
4:00 148.8 F 0.0
6:00 143.6 F 0.2
8:00 204.9 F 1.1
10:00 207.0 F 4.7
12:00 219.7 F 8.2
14:00 233.9 F 11.9
16:00 228.8 F 7.2
18:00 242.1 F 11.9
20:00 237.0 F 9.2
22:00 238.3 F 10.2
24:00 246.7 F 13.6
26:00 257.8 F 19.6
28:00 268.0 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.4
1:00 272.6 F 29.0
2:00 272.3 F 29.3
3:00 271.9 F 29.1
4:00 272.2 F 29.3

Min 270.4 F 28.3
Max 273.1 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.6 F 0.0
5:00 237.3 F 0.0
10:00 221.9 F 0.0
15:00 216.7 F 0.0
20:00 212.4 F 0.0
25:00 215.0 F 0.0
30:00 216.0 F 0.0
Drying Complete

FILLING: 0:36
HEATING: 29:13
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:06:58
CYCLE COMPLETE

10/12/2022 08:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____