

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3740

Name: RITTER 1  
Sterilizer ID: V2178198  
09/08/2023 10:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.6 F 0.0  
2:00 183.0 F 2.3  
4:00 203.8 F 7.0  
6:00 218.4 F 12.6  
8:00 226.9 F 7.8  
10:00 239.3 F 13.1  
12:00 239.4 F 10.7  
14:00 237.1 F 9.0  
16:00 250.8 F 15.6  
18:00 262.2 F 22.1  
20:00 272.7 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.7 F 28.8  
2:00 271.4 F 28.2  
3:00 272.7 F 28.9  
4:00 271.7 F 28.2  
  
Min 271.3 F 28.1  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.2 F 0.0  
5:00 204.8 F 0.0  
10:00 201.4 F 0.0  
15:00 220.2 F 0.0  
20:00 227.9 F 0.0  
25:00 224.7 F 0.0  
30:00 222.1 F 0.0  
Drying Complete

FILLING: 2:29  
HEATING: 20:24  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:00:16  
CYCLE COMPLETE

09/08/2023 11:12 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_