
Midmark M11 - v1.0.0
Total Cycles: 4519

Name: RITTER 1
Sterilizer ID: V2178198
05/02/2024 03:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.1 F 0.0
2:00 158.0 F 3.3
4:00 175.4 F 6.8
6:00 193.8 F 11.1
8:00 217.6 F 5.1
10:00 224.3 F 10.2
12:00 230.4 F 7.5
14:00 240.9 F 11.3
16:00 236.7 F 9.0
18:00 250.1 F 15.2
20:00 261.0 F 21.4
22:00 271.2 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 271.6 F 28.7
2:00 271.8 F 28.7
3:00 272.2 F 28.8
4:00 272.4 F 29.0

Min 271.4 F 28.6
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.3 F 0.0
5:00 225.0 F 0.0
10:00 201.5 F 0.0
15:00 195.5 F 0.0
20:00 222.7 F 0.0
25:00 236.3 F 0.0
30:00 195.1 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 23:23
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:02:23
CYCLE COMPLETE

05/02/2024 04:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____