
Midmark M11 - v1.0.0
Total Cycles: 3829

Name: RITTER 1
Sterilizer ID: V2178198
09/27/2023 01:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.1 F 0.0
2:00 179.5 F 0.9
4:00 200.1 F 5.3
6:00 220.8 F 10.5
8:00 227.5 F 6.3
10:00 239.5 F 11.8
12:00 237.6 F 9.3
14:00 238.4 F 9.1
16:00 249.7 F 14.7
18:00 262.1 F 21.7
20:00 272.4 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 27.9
1:00 272.1 F 28.1
2:00 272.5 F 28.6
3:00 271.5 F 27.8
4:00 272.5 F 28.5

Min 271.3 F 27.7
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.2 F 0.0
5:00 213.3 F 0.0
10:00 222.2 F 0.0
15:00 224.0 F 0.0
20:00 216.4 F 0.0
25:00 217.5 F 0.0
30:00 207.7 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 20:12
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:58:41
CYCLE COMPLETE

09/27/2023 02:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____