

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1433

Name: RITTER2  
Sterilizer ID: V2313228  
08/30/2022 11:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.9 F 0.0  
2:00 167.9 F 2.6  
4:00 195.0 F 7.3  
6:00 216.0 F 12.9  
8:00 225.9 F 8.9  
10:00 233.9 F 10.1  
12:00 241.7 F 12.8  
14:00 242.0 F 12.1  
16:00 255.3 F 19.0  
18:00 267.0 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.4  
1:00 272.6 F 30.1  
2:00 271.8 F 29.3  
3:00 272.2 F 30.1  
4:00 272.4 F 29.9  
  
Min 271.5 F 29.2  
Max 272.7 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.7 F 0.0  
5:00 213.2 F 0.0  
10:00 218.4 F 0.0  
15:00 226.6 F 0.0  
20:00 221.5 F 0.0  
25:00 216.5 F 0.0  
30:00 215.5 F 0.0  
Drying Complete

FILLING: 2:09  
HEATING: 19:29  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:50  
CYCLE COMPLETE

08/30/2022 12:08 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_