
Midmark M11 - v1.0.0
Total Cycles: 1894

Name: RITTER 1
Sterilizer ID: V2178198
05/02/2022 12:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.1 F 0.0
2:00 200.8 F 3.9
4:00 210.5 F 9.3
6:00 224.4 F 7.3
8:00 234.6 F 11.4
10:00 234.9 F 9.0
12:00 238.1 F 9.2
14:00 248.2 F 14.6
16:00 260.9 F 21.7
18:00 272.3 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.5 F 29.2
2:00 271.5 F 28.6
3:00 272.5 F 29.3
4:00 271.7 F 28.6

Min 271.4 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.5 F 0.0
5:00 212.6 F 0.0
10:00 229.9 F 0.0
15:00 232.9 F 0.0
20:00 226.7 F 0.0
25:00 231.2 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 3:33
HEATING: 18:28
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:09
CYCLE COMPLETE

05/02/2022 01:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____