
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006687
8:34 05/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:52
PRESSURIZING 3:53
STERILIZING 4:53
136.4 C 221kPa 4:53
135.9 C 220kPa 5:23
136.0 C 220kPa 5:53
136.4 C 223kPa 6:23
136.2 C 219kPa 6:53
136.4 C 222kPa 7:23
136.3 C 219kPa 7:53
136.4 C 222kPa 8:23
136.4 C 222kPa 8:23
VENTING 8:23
AIR DRYING 9:14
CYCLE COMPLETE 9:18

Digital Signature #
1596AC40E73DD470
