

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1606

Name: RITTER2  
Sterilizer ID: V2313228  
10/17/2022 10:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 200.6 F 0.0  
2:00 197.9 F 4.0  
4:00 199.8 F 9.0  
6:00 220.5 F 8.4  
8:00 230.1 F 10.6  
10:00 232.0 F 8.2  
12:00 240.2 F 11.0  
14:00 246.0 F 13.9  
16:00 258.3 F 20.6  
18:00 269.2 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.3  
1:00 272.4 F 30.1  
2:00 272.3 F 29.6  
3:00 271.4 F 29.1  
4:00 271.5 F 29.6  
  
Min 271.2 F 29.1  
Max 272.9 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.7 F 0.0  
5:00 224.0 F 0.0  
10:00 242.1 F 0.0  
15:00 247.5 F 0.0  
20:00 235.0 F 0.0  
25:00 226.2 F 0.0  
30:00 224.4 F 0.0  
Drying Complete

FILLING: 2:21  
HEATING: 19:10  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:45  
CYCLE COMPLETE

10/17/2022 11:48 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_