
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002466
15:26 28/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:45
PRESSURIZING 2:51
STERILIZING 4:03
136.0 C 220kPa 4:03
136.4 C 225kPa 4:33
136.3 C 225kPa 5:03
136.1 C 223kPa 5:33
136.4 C 223kPa 6:03
136.2 C 222kPa 6:33
136.4 C 223kPa 7:03
136.4 C 226kPa 7:33
136.4 C 226kPa 7:33
VENTING 7:33
AIR DRYING 8:28
CYCLE COMPLETE 24:25

Digital Signature #
AB49CE408BF1F03C
