
Midmark M11 - v1.0.5
Total Cycles: 1276

Name: RITTER2
Sterilizer ID: V2313228
07/08/2022 02:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.2 F 0.0
2:00 169.0 F 3.0
4:00 189.2 F 7.7
6:00 208.8 F 13.1
8:00 224.5 F 7.9
10:00 237.4 F 13.4
12:00 238.7 F 11.2
14:00 237.7 F 10.0
16:00 251.6 F 16.8
18:00 263.2 F 23.6
20:00 272.2 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.2
1:00 272.6 F 29.7
2:00 271.5 F 29.1
3:00 272.6 F 29.9
4:00 272.1 F 29.3

Min 271.5 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 210.6 F 0.0
10:00 224.4 F 0.0
15:00 228.3 F 0.0
20:00 222.3 F 0.0
25:00 221.7 F 0.0
30:00 218.9 F 0.0
Drying Complete

FILLING: 2:40
HEATING: 20:09
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:58
CYCLE COMPLETE

07/08/2022 03:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____