
Midmark M11 - v1.0.5
Total Cycles: 334

Name: RITTER2
Sterilizer ID: V2313228
09/02/2021 03:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.7 F 0.0
2:00 177.1 F 3.6
4:00 201.1 F 8.6
6:00 220.7 F 8.6
8:00 231.0 F 10.6
10:00 233.1 F 8.5
12:00 239.9 F 10.5
14:00 247.7 F 14.8
16:00 260.1 F 21.6
18:00 271.4 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.2 F 29.5
2:00 271.5 F 29.5
3:00 272.1 F 29.4
4:00 271.4 F 29.5

Min 271.1 F 29.0
Max 272.5 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.3 F 0.0
5:00 211.8 F 0.0
10:00 209.0 F 0.0
15:00 208.2 F 0.0
20:00 203.7 F 0.0
25:00 200.8 F 0.0
30:00 198.7 F 0.0
Drying Complete

FILLING: 3:39
HEATING: 18:36
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:24
CYCLE COMPLETE

09/02/2021 04:44 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____