
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003180
10:47 17/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:38
STERILIZING 3:38
136.4 C 220kPa 3:38
136.4 C 225kPa 4:08
136.3 C 222kPa 4:38
136.2 C 220kPa 5:08
135.9 C 221kPa 5:38
136.4 C 225kPa 6:08
136.4 C 224kPa 6:38
136.2 C 223kPa 7:08
136.2 C 223kPa 7:08
VENTING 7:08
AIR DRYING 7:58
CYCLE COMPLETE 8:08

Digital Signature #
79854F64313812ED
