
Midmark M11 - v1.0.5
Total Cycles: 3007

Name: RITTER2
Sterilizer ID: V2313228
10/13/2023 10:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.0 F 0.0
2:00 186.3 F 3.5
4:00 206.3 F 9.3
6:00 223.9 F 7.1
8:00 234.7 F 11.2
10:00 235.4 F 9.3
12:00 237.4 F 9.2
14:00 250.6 F 16.0
16:00 263.3 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.3 F 29.3
2:00 272.0 F 29.5
3:00 272.2 F 29.2
4:00 271.8 F 29.4

Min 271.5 F 28.8
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.9 F 0.0
5:00 225.9 F 0.0
10:00 226.2 F 0.0
15:00 222.4 F 0.0
20:00 219.4 F 0.0
25:00 218.7 F 0.0
30:00 217.1 F 0.0
Drying Complete

FILLING: 2:03
HEATING: 17:60
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:57:18
CYCLE COMPLETE

10/13/2023 11:54 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____