
Midmark M11 - v1.0.5
Total Cycles: 556

Name: RITTER2
Sterilizer ID: V2313228
11/17/2021 10:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 106.4 F 0.0
2:00 106.4 F 0.0
4:00 120.9 F 0.3
6:00 155.3 F 3.7
8:00 179.4 F 7.6
10:00 199.8 F 12.6
12:00 212.1 F 6.3
14:00 226.0 F 10.9
16:00 230.1 F 7.5
18:00 242.7 F 13.2
20:00 242.8 F 12.4
22:00 255.4 F 18.8
24:00 266.2 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.4 F 29.7
2:00 271.3 F 29.5
3:00 272.3 F 29.7
4:00 271.2 F 29.2

Min 271.2 F 29.1
Max 272.5 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.0 F 0.0
5:00 222.7 F 0.0
10:00 223.7 F 0.0
15:00 214.2 F 0.0
20:00 204.1 F 0.0
25:00 197.9 F 0.0
30:00 205.2 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 25:37
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:03:40
CYCLE COMPLETE

11/17/2021 11:17 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____