
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006259
9:06 07/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 2:00
STERILIZING 3:12
136.4 C 223kPa 3:12
136.2 C 223kPa 3:42
136.2 C 221kPa 4:12
136.2 C 222kPa 4:42
136.1 C 222kPa 5:12
136.1 C 220kPa 5:42
136.2 C 219kPa 6:12
136.1 C 219kPa 6:42
136.1 C 219kPa 6:42
VENTING 6:42
AIR DRYING 7:29
CYCLE COMPLETE 7:43

Digital Signature #
C98E222050F7507C
