

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2739

Name: RITTER2  
Sterilizer ID: V2313228  
08/08/2023 12:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 152.9 F 0.0  
2:00 155.9 F 0.3  
4:00 164.4 F 3.7  
6:00 187.3 F 8.2  
8:00 209.1 F 12.1  
10:00 219.3 F 8.2  
12:00 232.3 F 13.5  
14:00 237.7 F 10.8  
16:00 235.1 F 9.0  
18:00 249.2 F 15.5  
20:00 260.7 F 21.9  
22:00 270.9 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.1  
1:00 271.9 F 29.3  
2:00 271.9 F 29.8  
3:00 271.3 F 29.0  
4:00 271.9 F 29.6

Min 271.0 F 28.9  
Max 272.3 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.4 F 0.0  
5:00 208.0 F 0.0  
10:00 211.7 F 0.0  
15:00 206.6 F 0.0  
20:00 201.0 F 0.0  
25:00 212.8 F 0.0  
30:00 213.0 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 22:41  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:00:48  
CYCLE COMPLETE

08/08/2023 01:12 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_