

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2562

Name: RITTER2  
Sterilizer ID: V2313228  
06/23/2023 03:47 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 204.9 F 0.0  
2:00 198.6 F 4.3  
4:00 212.5 F 10.0  
6:00 225.7 F 7.0  
8:00 237.6 F 12.3  
10:00 238.1 F 10.6  
12:00 237.4 F 9.7  
14:00 251.9 F 16.7  
16:00 263.9 F 23.7  
18:00 271.4 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.6 F 29.5  
2:00 271.5 F 28.8  
3:00 272.4 F 29.8  
4:00 272.2 F 29.3  
  
Min 271.3 F 28.8  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.2 F 0.0  
5:00 226.6 F 0.0  
10:00 229.0 F 0.0  
15:00 227.0 F 0.0  
20:00 223.2 F 0.0  
25:00 224.0 F 0.0  
30:00 223.4 F 0.0  
Drying Complete

FILLING: 2:54  
HEATING: 18:03  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:11  
CYCLE COMPLETE

06/23/2023 04:45 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_