

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1215

Name: RITTER 1  
Sterilizer ID: V2178198  
09/29/2021 06:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.1 F 0.0  
2:00 77.2 F 0.0  
4:00 87.4 F 0.0  
6:00 103.2 F 0.1  
8:00 132.8 F 0.1  
10:00 162.0 F 0.2  
12:00 184.9 F 2.0  
14:00 204.2 F 4.6  
16:00 212.4 F 7.7  
18:00 223.9 F 11.2  
20:00 225.3 F 6.2  
22:00 235.6 F 10.4  
24:00 235.5 F 8.8  
26:00 244.3 F 12.7  
28:00 242.1 F 11.2  
30:00 253.9 F 17.2  
32:00 264.3 F 23.3  
34:00 271.9 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.1  
1:00 271.9 F 28.6  
2:00 272.3 F 28.6  
3:00 271.4 F 28.0  
4:00 272.5 F 28.9  
  
Min 271.4 F 28.0  
Max 272.6 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.9 F 0.0  
5:00 202.8 F 0.0  
10:00 209.1 F 0.0  
15:00 205.7 F 0.0  
20:00 204.2 F 0.0  
25:00 198.5 F 0.0  
30:00 201.2 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 34:01  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:12:14  
CYCLE COMPLETE

09/29/2021 07:46 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01