

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3819

Name: RITTER2  
Sterilizer ID: V2313228  
07/26/2024 02:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.8 F 0.0  
2:00 160.9 F 3.0  
4:00 178.8 F 7.2  
6:00 206.6 F 12.3  
8:00 218.8 F 7.5  
10:00 232.8 F 12.9  
12:00 235.3 F 10.3  
14:00 236.5 F 8.6  
16:00 249.2 F 15.6  
18:00 261.3 F 22.4  
20:00 272.1 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.4  
1:00 273.0 F 30.1  
2:00 271.5 F 29.4  
3:00 271.7 F 30.0  
4:00 272.7 F 29.9

Min 271.1 F 29.2  
Max 273.2 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.9 F 0.0  
5:00 209.6 F 0.0  
10:00 219.1 F 0.0  
15:00 216.2 F 0.0  
20:00 214.0 F 0.0  
25:00 213.9 F 0.0  
30:00 213.3 F 0.0  
Drying Complete

FILLING: 1:29  
HEATING: 20:33  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 00:59:29  
CYCLE COMPLETE

07/26/2024 03:42 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_