

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1155

Name: RITTER2  
Sterilizer ID: V2313228  
06/01/2022 04:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.0 F 0.0  
2:00 185.1 F 3.9  
4:00 199.1 F 9.2  
6:00 216.5 F 7.1  
8:00 231.7 F 10.3  
10:00 233.3 F 8.0  
12:00 241.8 F 11.3  
14:00 247.1 F 14.0  
16:00 259.9 F 21.1  
18:00 271.5 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.4 F 29.0  
2:00 272.1 F 29.3  
3:00 272.0 F 28.8  
4:00 272.3 F 29.4  
  
Min 271.4 F 28.5  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.8 F 0.0  
5:00 208.1 F 0.0  
10:00 226.2 F 0.0  
15:00 218.9 F 0.0  
20:00 217.2 F 0.0  
25:00 216.6 F 0.0  
30:00 220.1 F 0.0  
Drying Complete

FILLING: 2:01  
HEATING: 18:35  
STERILIZING: 4:00  
VENTING: 3:02  
DRYING: 30:00  
TOTAL CYCLE: 00:57:39  
CYCLE COMPLETE

06/01/2022 05:12 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_