

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001432  
13:16 14/08/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:13  
PRESSURIZING 3:15  
STERILIZING 4:22  
136.4 C 224kPa 4:22  
136.4 C 226kPa 4:52  
136.2 C 224kPa 5:22  
136.3 C 225kPa 5:52  
136.2 C 224kPa 6:22  
136.2 C 223kPa 6:52  
136.2 C 224kPa 7:22  
136.4 C 225kPa 7:52  
136.4 C 225kPa 7:52  
VENTING 7:52  
AIR DRYING 8:46  
CYCLE COMPLETE 11:20

-----  
Digital Signature #  
2EDF293C5F03E46D  
-----