
Midmark M11 - v1.0.0
Total Cycles: 767

Name: RITTER 1
Sterilizer ID: V2178198
04/01/2021 01:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.6 F 0.0
2:00 153.9 F 0.3
4:00 157.8 F 2.7
6:00 185.0 F 6.4
8:00 203.3 F 11.3
10:00 219.6 F 5.3
12:00 229.3 F 10.5
14:00 230.1 F 7.2
16:00 244.7 F 13.3
18:00 243.1 F 11.6
20:00 256.1 F 18.5
22:00 267.3 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 271.3 F 28.4
2:00 271.5 F 28.5
3:00 272.0 F 28.8
4:00 271.9 F 29.1

Min 271.1 F 28.4
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.0 F 0.0
5:00 230.9 F 0.0
10:00 207.5 F 0.0
15:00 231.2 F 0.0
20:00 211.2 F 0.0
25:00 181.9 F 0.0
Dry Cycle Aborted

FILLING: 1:08
HEATING: 23:13
STERILIZING: 4:00
VENTING: 3:01
DRYING: 28:29
TOTAL CYCLE: 00:59:52

04/01/2021 02:25 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____