
Midmark M11 - v1.0.5
Total Cycles: 3628

Name: RITTER2
Sterilizer ID: V2313228
05/17/2024 10:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.0 F 0.0
2:00 71.5 F 0.0
4:00 97.5 F 0.0
6:00 142.6 F 0.5
8:00 173.6 F 2.6
10:00 190.8 F 6.2
12:00 196.5 F 10.6
14:00 214.0 F 4.8
16:00 225.6 F 7.4
18:00 236.3 F 11.7
20:00 233.2 F 8.2
22:00 243.2 F 11.9
24:00 244.2 F 12.6
26:00 257.0 F 19.4
28:00 268.1 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.0 F 28.7
2:00 272.8 F 29.2
3:00 272.2 F 29.3
4:00 271.6 F 28.5

Min 271.5 F 28.5
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.0 F 0.0
5:00 209.6 F 0.0
10:00 221.6 F 0.0
15:00 220.4 F 0.0
20:00 215.1 F 0.0
25:00 214.0 F 0.0
30:00 213.9 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 29:12
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 01:07:02
CYCLE COMPLETE

05/17/2024 11:10 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____