

-----  
Midmark M11 - v1.0.5  
Total Cycles: 335

Name: RITTER2  
Sterilizer ID: V2313228  
09/02/2021 04:47 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.3 F 0.0  
2:00 184.6 F 3.8  
4:00 202.5 F 9.3  
6:00 220.3 F 7.7  
8:00 232.8 F 11.3  
10:00 235.5 F 9.8  
12:00 238.1 F 9.2  
14:00 251.1 F 16.6  
16:00 263.5 F 23.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.1  
1:00 272.7 F 29.6  
2:00 272.2 F 29.7  
3:00 272.3 F 29.4  
4:00 272.2 F 29.8  
  
Min 271.6 F 29.0  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.9 F 0.0  
5:00 207.8 F 0.0  
10:00 211.6 F 0.0  
15:00 207.8 F 0.0  
Dry Cycle Aborted

FILLING: 3:52  
HEATING: 17:59  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 17:07  
TOTAL CYCLE: 00:46:10

09/02/2021 05:33 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02