
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004183
14:41 09/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:56
PRESSURIZING 3:00
STERILIZING 4:06
136.4 C 223kPa 4:06
136.4 C 225kPa 4:36
136.3 C 223kPa 5:06
136.4 C 225kPa 5:36
136.3 C 224kPa 6:06
135.9 C 221kPa 6:36
136.4 C 223kPa 7:06
136.2 C 222kPa 7:36
136.2 C 222kPa 7:36
VENTING 7:36
AIR DRYING 8:29
CYCLE COMPLETE 8:36

Digital Signature #
02EEB91D31E8C6FF
