
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000237
8:34 29/09/2020
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CYCLE ABORTED 0:03
STOP BUTTON PRESSED
NOT STERILE
END TIME 0:03

Digital Signature #
B3E3FFDD82A36EE5
