
Midmark M11 - v1.0.5
Total Cycles: 4086

Name: RITTER2
Sterilizer ID: V2313228
11/05/2024 12:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.6 F 0.0
2:00 158.8 F 2.1
4:00 165.2 F 5.4
6:00 189.8 F 9.5
8:00 205.8 F 10.9
10:00 220.6 F 7.9
12:00 231.9 F 12.4
14:00 233.0 F 9.3
16:00 238.5 F 10.4
18:00 244.6 F 13.3
20:00 255.9 F 19.2
22:00 265.9 F 25.5
24:00 271.6 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 29.7
1:00 272.7 F 30.3
2:00 271.7 F 29.7
3:00 272.8 F 30.1
4:00 271.5 F 29.8

Min 271.5 F 29.6
Max 272.9 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 212.5 F 0.0
10:00 206.2 F 0.0
15:00 210.4 F 0.0
20:00 213.6 F 0.0
25:00 216.0 F 0.0
30:00 218.1 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 24:27
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:03:30
CYCLE COMPLETE

11/05/2024 02:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____