

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
1.0uS / 0.6ppm  
CYCLE NUMBER 001551  
12:17 11/03/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 5:11  
PRESSURIZING 6:56  
STERILIZING 8:34  
136.0 C 220kPa 8:34  
136.4 C 222kPa 9:04  
136.2 C 221kPa 9:34  
135.8 C 217kPa 10:04  
136.5 C 222kPa 10:34  
136.5 C 223kPa 11:04  
136.5 C 220kPa 11:34  
136.3 C 221kPa 12:04  
VENTING 12:04  
AIR DRYING 13:08  
CYCLE COMPLETE 32:11

-----  
Digital Signature #  
775BE246752BCCE9  
-----