
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004301
11:34 06/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:26
PRESSURIZING 1:29
STERILIZING 2:30
136.4 C 220kPa 2:30
136.4 C 225kPa 3:00
136.4 C 222kPa 3:30
136.2 C 223kPa 4:00
136.4 C 222kPa 4:30
136.2 C 222kPa 5:00
136.4 C 225kPa 5:30
136.2 C 221kPa 6:00
136.2 C 221kPa 6:00
VENTING 6:00
AIR DRYING 6:47
CYCLE COMPLETE 7:11

Digital Signature #
5E9490621FF085D8
