
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 006010
11:21 19/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:24
PRESSURIZING 2:30
STERILIZING 3:28
136.1 C 221kPa 3:28
135.9 C 218kPa 3:58
136.1 C 221kPa 4:28
136.2 C 222kPa 4:58
136.4 C 222kPa 5:28
136.4 C 223kPa 5:58
136.4 C 223kPa 6:28
136.2 C 222kPa 6:58
136.2 C 222kPa 6:58
VENTING 6:58
AIR DRYING 7:51
CYCLE COMPLETE 8:23

Digital Signature #
CCCDF73812242C4C
