
Midmark M11 - v1.0.5
Total Cycles: 1043

Name: RITTER2
Sterilizer ID: V2313228
04/27/2022 12:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.1 F 0.0
2:00 178.3 F 3.6
4:00 189.5 F 8.0
6:00 208.9 F 13.4
8:00 222.7 F 8.2
10:00 235.3 F 11.9
12:00 239.5 F 11.8
14:00 239.8 F 10.9
16:00 253.2 F 17.7
18:00 265.1 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.4 F 29.8
2:00 271.5 F 29.3
3:00 272.5 F 30.0
4:00 272.0 F 29.4

Min 271.5 F 29.1
Max 272.5 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.9 F 0.0
5:00 218.8 F 0.0
10:00 234.2 F 0.0
15:00 229.1 F 0.0
20:00 227.0 F 0.0
25:00 226.9 F 0.0
30:00 224.8 F 0.0

Drying Complete

FILLING: 1:28
HEATING: 19:49
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:28
CYCLE COMPLETE

04/27/2022 12:58 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____