
Midmark M11 - v1.0.5
Total Cycles: 1841

Name: RITTER2
Sterilizer ID: V2313228
12/14/2022 02:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.3 F 0.0
2:00 189.2 F 4.4
4:00 198.0 F 9.5
6:00 217.5 F 6.1
8:00 232.6 F 10.2
10:00 234.3 F 8.1
12:00 241.5 F 10.9
14:00 248.5 F 14.6
16:00 261.8 F 22.0
18:00 272.6 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.4 F 29.0
2:00 271.9 F 29.1
3:00 272.5 F 29.0
4:00 271.5 F 28.7

Min 271.5 F 28.4
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.6 F 0.0
5:00 215.0 F 0.0
10:00 232.5 F 0.0
15:00 220.2 F 0.0
20:00 216.3 F 0.0
25:00 222.8 F 0.0
30:00 223.0 F 0.0

Drying Complete

FILLING: 1:51
HEATING: 18:15
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:16
CYCLE COMPLETE

12/14/2022 03:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____