
Midmark M11 - v1.0.0
Total Cycles: 3821

Name: RITTER 1
Sterilizer ID: V2178198
09/26/2023 01:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.5 F 0.0
2:00 175.3 F 2.8
4:00 199.4 F 7.5
6:00 215.0 F 12.8
8:00 224.7 F 7.7
10:00 236.6 F 12.9
12:00 238.6 F 10.4
14:00 236.4 F 8.5
16:00 250.5 F 15.4
18:00 262.1 F 21.9
20:00 272.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 271.9 F 28.3
2:00 272.3 F 28.7
3:00 271.3 F 28.3
4:00 271.8 F 28.3

Min 271.1 F 28.0
Max 272.4 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.2 F 0.0
5:00 212.4 F 0.0
10:00 230.5 F 0.0
15:00 231.1 F 0.0
20:00 225.6 F 0.0
25:00 222.7 F 0.0
30:00 223.1 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 20:20
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 00:59:26
CYCLE COMPLETE

09/26/2023 02:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____