
Midmark M11 - v1.0.0
Total Cycles: 4319

Name: RITTER 1
Sterilizer ID: V2178198
02/21/2024 06:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 117.8 F 0.0
2:00 117.7 F 1.1
4:00 124.6 F 3.6
6:00 154.9 F 6.7
8:00 177.3 F 10.5
10:00 212.0 F 5.9
12:00 220.8 F 7.7
14:00 229.9 F 11.8
16:00 230.4 F 7.5
18:00 241.4 F 12.5
20:00 239.0 F 10.3
22:00 250.6 F 16.1
24:00 261.4 F 22.1
26:00 271.3 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.3 F 29.3
2:00 272.7 F 29.6
3:00 271.6 F 29.3
4:00 272.1 F 29.2

Min 271.4 F 29.0
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.8 F 0.0
5:00 225.2 F 0.0
10:00 233.4 F 0.0
15:00 230.2 F 0.0
20:00 228.6 F 0.0
25:00 227.4 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 27:22
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:06:25
CYCLE COMPLETE

02/21/2024 07:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____