
Midmark M11 - v1.0.0
Total Cycles: 5442

Name: RITTER 1
Sterilizer ID: V2178198
03/25/2025 11:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.2 F 0.0
2:00 145.1 F 2.0
4:00 156.8 F 4.9
6:00 180.0 F 8.4
8:00 198.7 F 11.9
10:00 208.6 F 4.7
12:00 218.6 F 9.3
14:00 238.1 F 10.7
16:00 236.9 F 10.0
18:00 238.1 F 9.3
20:00 245.0 F 13.3
22:00 256.4 F 19.1
24:00 265.9 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.0 F 29.7
2:00 272.4 F 29.5
3:00 271.6 F 29.0
4:00 271.0 F 29.0

Min 271.0 F 28.9
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.5 F 0.0
5:00 205.5 F 0.0
10:00 208.7 F 0.0
15:00 211.8 F 0.0
20:00 218.7 F 0.0
25:00 224.9 F 0.0
30:00 213.3 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 25:50
STERILIZING: 4:00
VENTING: 3:60
DRYING: 30:00
TOTAL CYCLE: 01:04:41
CYCLE COMPLETE

03/25/2025 12:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____