

-----  
Midmark M11 - v1.0.5  
Total Cycles: 352

Name: RITTER2  
Sterilizer ID: V2313228  
09/13/2021 06:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.3 F 0.0  
2:00 71.9 F 0.0  
4:00 76.5 F 0.1  
6:00 100.0 F 0.2  
8:00 126.8 F 2.1  
10:00 153.0 F 4.5  
12:00 179.5 F 7.0  
14:00 194.0 F 10.7  
16:00 211.5 F 6.5  
18:00 223.3 F 7.2  
20:00 232.9 F 10.9  
22:00 231.8 F 8.0  
24:00 243.2 F 12.5  
26:00 240.1 F 11.0  
28:00 251.6 F 16.7  
30:00 261.5 F 22.4  
32:00 270.7 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 29.1  
1:00 272.7 F 29.8  
2:00 272.1 F 29.0  
3:00 272.7 F 29.9  
4:00 271.9 F 29.2  
  
Min 271.7 F 28.9  
Max 273.3 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.1 F 0.0  
5:00 202.9 F 0.0  
10:00 192.6 F 0.0  
15:00 195.5 F 0.0  
20:00 195.0 F 0.0  
25:00 192.7 F 0.0  
30:00 198.8 F 0.0  
  
Drying Complete

FILLING: 0:56  
HEATING: 32:41  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:10:57  
CYCLE COMPLETE

09/13/2021 07:58 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01