
Midmark M11 - v1.0.0
Total Cycles: 2957

Name: RITTER 1
Sterilizer ID: V2178198
02/28/2023 05:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.6 F 0.0
2:00 184.8 F 3.1
4:00 194.2 F 6.9
6:00 221.0 F 11.7
8:00 218.6 F 6.2
10:00 231.8 F 11.4
12:00 233.4 F 8.4
14:00 239.9 F 10.4
16:00 245.8 F 13.5
18:00 258.8 F 20.5
20:00 270.4 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.3 F 29.5
2:00 271.7 F 28.8
3:00 272.0 F 29.4
4:00 272.3 F 29.2

Min 271.3 F 28.6
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.6 F 0.0
5:00 228.1 F 0.0
10:00 235.6 F 0.0
15:00 233.8 F 0.0
20:00 226.1 F 0.0
25:00 226.5 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 20:41
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:01
CYCLE COMPLETE

02/28/2023 05:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____