
Midmark M11 - v1.0.5
Total Cycles: 2702

Name: RITTER2
Sterilizer ID: V2313228
07/26/2023 07:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.5 F 0.0
2:00 81.3 F 0.0
4:00 121.2 F 0.0
6:00 134.6 F 0.0
8:00 191.1 F 0.0
10:00 176.6 F 0.4
12:00 193.4 F 2.9
14:00 218.3 F 5.7
16:00 225.5 F 8.7
18:00 233.6 F 12.1
20:00 229.6 F 7.1
22:00 239.4 F 11.7
24:00 236.0 F 9.0
26:00 240.9 F 10.7
28:00 246.2 F 13.5
30:00 257.3 F 19.4
32:00 267.4 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.6 F 29.1
2:00 271.4 F 28.7
3:00 272.5 F 29.2
4:00 271.4 F 28.5

Min 271.3 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.9 F 0.0
5:00 222.5 F 0.0
10:00 232.5 F 0.0
15:00 226.7 F 0.0
20:00 223.1 F 0.0
25:00 229.7 F 0.0
30:00 229.8 F 0.0

Drying Complete

FILLING: 0:48
HEATING: 33:31
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:11:36
CYCLE COMPLETE

07/26/2023 09:08 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____