
Midmark M11 - v1.0.0
Total Cycles: 3194

Name: RITTER 1
Sterilizer ID: V2178198
05/03/2023 10:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.9 F 0.0
2:00 193.4 F 3.2
4:00 216.9 F 8.9
6:00 226.7 F 7.6
8:00 234.5 F 11.0
10:00 234.6 F 8.7
12:00 238.6 F 9.7
14:00 248.8 F 14.8
16:00 261.6 F 22.0
18:00 272.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.5 F 29.1
2:00 271.3 F 28.4
3:00 272.8 F 29.3
4:00 271.8 F 28.6

Min 271.2 F 28.4
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.5 F 0.0
5:00 231.5 F 0.0
10:00 238.8 F 0.0
15:00 230.3 F 0.0
20:00 219.1 F 0.0
25:00 211.0 F 0.0
30:00 208.4 F 0.0

Drying Complete

FILLING: 2:03
HEATING: 18:20
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:57:40
CYCLE COMPLETE

05/03/2023 11:52 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____