
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004737
13:30 21/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:24
PRESSURIZING 2:26
STERILIZING 3:16
136.0 C 220kPa 3:16
136.4 C 225kPa 3:46
136.4 C 222kPa 4:16
136.3 C 224kPa 4:46
136.1 C 223kPa 5:16
136.4 C 223kPa 5:46
136.4 C 223kPa 6:16
136.4 C 222kPa 6:46
136.4 C 222kPa 6:46
VENTING 6:46
AIR DRYING 7:36
CYCLE COMPLETE 7:45

Digital Signature #
BBDED6FA52AE10D4
