
Midmark M11 - v1.0.0
Total Cycles: 4086

Name: RITTER 1
Sterilizer ID: V2178198
12/05/2023 03:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.3 F 0.0
2:00 143.9 F 1.9
4:00 152.5 F 5.0
6:00 188.2 F 8.7
8:00 202.4 F 13.5
10:00 214.8 F 5.8
12:00 225.3 F 10.1
14:00 232.6 F 8.5
16:00 239.7 F 11.3
18:00 237.6 F 9.4
20:00 251.0 F 15.7
22:00 262.1 F 22.1
24:00 272.3 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.5 F 29.2
2:00 272.1 F 28.8
3:00 271.2 F 28.3
4:00 271.5 F 28.8

Min 271.0 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.0 F 0.0
5:00 218.4 F 0.0
10:00 240.1 F 0.0
15:00 233.0 F 0.0
20:00 225.7 F 0.0
25:00 224.4 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 24:32
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:03:24
CYCLE COMPLETE

12/05/2023 04:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____